

DECLARATION OF CONFORMITY

Manufacturer's Name: Maxtec, LLC
Address: 2305 South 1070 West
Salt Lake City, Utah 84119
USA

SRN: US-MF-000003961

European Representative: EMERGO EUROPE
Prinsessegracht 20
2515 AP The Hague
The Netherlands

Basic UDI DI: Pending Assignment

Product Name(s): UltraMaxO2

Intended Purpose: The UltraMaxO2 Oxygen Analyzer is a tool intended to measure oxygen purity, flow and pressure of an oxygen concentrator.

Risk Class & GMDN: Class IIa 46049 – Respiratory Oxygen Monitor, Battery Powered

Classification Criteria: Clause 3.2 Rule 10 of Annex IX of MDD

This EU declaration of conformity is issued under the sole responsibility of the manufacturer. We herewith declare that the above mentioned products meet the provisions of the following EC Council Regulations, Common Specifications and Standards. All supporting documents are retained under the premises of the manufacturer and the European Representative.

Regulations: General Application Directives: Medical Device Directive, Council Directive 93/42/EEC of 14 June 1993 per Annex II Excluding Section 4 as amended by 2007/47/EC of September 2007

Notified Body: TÜV SÜD Product Service
RIDLERSTRASSE 65, D-80339 MUNICH, Germany
Number 0123

Conformity Assessment: Full Quality Assurance per Annex II MDD 93/42/EEC Excluding Section 4

EC Certificate No.: G1 045041 00 25 Rev. 00



Date CE mark was affixed: 15 December 2011

This declaration is considered valid from date of signature below until May 26, 2024

Signature: _____ Date: 5/15/2021

Name: Sidra Hankins
Position: VP of QA/RA

Applied Common Specification and Standards

The referenced list of Common Specifications and harmonized standards for which documented evidence of compliance can be provided includes:

- EN ISO 13485:2016 (ISO 13485:2016)
- EN ISO 14971:2012 (ISO 14971:2007, Corrected version 2007-10-01)
- EN 1041:2008
- EN ISO 15223-1:2016 (ISO 15223-1:2016, Corrected version 2017-03)



SENSING. ANALYSIS. DELIVERY.

September 22, 2021

Distributed by: Invacare – Manufactured by: Maxtec, LLC
Check O2 Plus - UltraMaxO2

To whom it may concern,

Invacare's, located at One Invacare Way, Elyria, Ohio, 44035, USA, Check O2 Plus, Reference Number: IRC450, is a distributed labeled variant of the UltraMaxO2.

A handwritten signature in black ink, appearing to read "Harris Baniissa", with a long horizontal flourish extending to the right.

Harris Baniissa
RA/QA Product Engineering Specialist
September 22, 2021

The Maxtec logo, identical to the one at the top left of the page, is positioned to the left of the company's address.

maxtec
2305 South 1070 West
Salt Lake City, UT 84119

WEB: www.maxtec.com

PHONE: 866.4.Maxtec

2305 S. 1070 W. • SLC, Utah • 84119 • USA