



TÜV SÜD Product Service GmbH • Masurenweg 1-3 • 30163 Hanover • Germany

Invacare EC-Hong A/S  
Mr Jan Kristoffersen  
Ostergade 3  
4270 Hong  
DENMARK

Your ref./from	Our ref./Name	Phone-ext/E-Mail	Fax-ext	Date	Page
	PS-MHS5-HAN-zt/ahe Torsten Zimmer	+49 511 9663-823 torsten.zimmer@tuev-sued.de	+49 511 9663-839	2011-03-04	1 of 1

### Side-rail type "Britt 5"

Dear Mr Kristoffersen,

Please find enclosed our Technical Report No. 71383165 about testing the above mentioned product.

Please advice how to proceed with the test sample.

In case of further questions please do not hesitate to contact directly Mr Dirk Wexlberger, phone: +49/511/9663-863.

Yours sincerely,

TÜV SÜD Product Service GmbH

i.A. Torsten Zimmer, Dipl.-Ing.  
Department Rehabilitation

encl.  
report  
invoice

Trade Register Munich HRB 85742  
HypoVereinsbank Munich  
Acc. No. 48 852 211  
Bank Code 700 202 70  
VAT ID No. DE129484267  
Information pursuant to Section 2(1)  
DL-InfoV (Germany)  
at [www.tuev-sued.com/imprint](http://www.tuev-sued.com/imprint)

Supervisory Board:  
Dirk Eilers (Chairman)  
  
Management:  
Joachim Birnthalder (CEO)  
Dr. Jens Butenandt (Managing Director)

Telefon: +49 511 9663-6  
Telefax: +49 511 9663-779

[www.tuev-sued.de](http://www.tuev-sued.de)  
TUV®

TÜV SÜD Product Service GmbH  
  
Hanover Branch  
Masurenweg 1-3  
30163 Hanover  
Germany



Product Service

Choose certainty.  
Add value.

## Technical Report No. 71383165

Revision: 0  
dated 2011-03-04

Client: Invacare EC-Hong A/S  
Ostergade 3  
4270 HONG  
DENMARK

Manufacturing place: Invacare EC-Hong A/S  
Ostergade 3  
4270 HONG  
DENMARK

Test object: Side rail for medical beds  
- type "BRITT 5"

Test specifications: IEC 60601-2-52:2009, section 201.9.8.3.3.3

Purpose of examination: Testing according to the test specifications.

Test result: The test subject was found to be in compliance with the test specifications.

This technical report may only be quoted in full. Any use for advertising purposes must be granted in writing. This report is the result of a single examination of the object in question and is not generally applicable evaluation of the quality of other products in regular production.