

# Test report No.: PB - 13 - 122 - MP - PA 138 - E




DIN EN 60601-2-52: 2010 Medical electrical equipment -  
Part 2-52: Particular requirements for basic safety and essential  
performance of medical beds (IEC 60601-2-52:2009)  
Page 1/30, 2013-09-19

**Test institute:** Berlin Cert - Prüf- und Zertifizierstelle für Medizinprodukte GmbH  
an der Technischen Universität Berlin  
Dovestr. 6, D-10587 Berlin  
Tel.: +49(0)30-314-25111, Fax: -23719

**Client:** Invacare Rea AB, Box 200, Växjövägen 303, SCHWEDEN-34375 Diö

**Manufacturer:** Invacare Rea AB, Box 200, Växjövägen 303, SCHWEDEN-34375 Diö

**Test sample:** Type / Model: Care bed Type: 1557011-0152 / Medley Ergo  
Quantity / Identification: 1 / S/N: 13F09000635  
Receipt / Condition: 2013-06-12 / new



**Test standard:** DIN EN 60601-2-52: 2010 Medical electrical equipment -  
Part 2-52: Particular requirements for basic safety and essential performance  
of medical beds (IEC 60601-2-52:2009)

**Annotation:** The numbering corresponds to the applied standard DIN EN 60601-2-52:  
2010, (IEC 60601-2-52:2009)

**Accreditation:** Accredited by "Deutsche Akkreditierungsstelle GmbH (DAKkS)"  
(Reg.no.: D-PL-17591-01-01)

**Kind of the test:** Complete test


**Test period:** 2013-06-18 to 2013-09-05


**Test location:** Rooms of the test institute

**Test result:** The test requirements are fulfilled.

established: 2013-09-19

released: 2013-09-19

  
Dipl.-Ing. (FH) M. Fiedler  
Test Engineer

  
Dipl.-Ing. (FH) F. Abmann  
Technical Director (Test Laboratory)

This test report may only be quoted in full. Any use for advertising purpose must be granted in writing.  
This report is the result of a single examination of the object in question and is not generally applicable evaluation of the other products in regular production.