

Certificate

No.: Z1 02 11 42940 009

TÜV
PRODUCT SERVICE

Invacare EC-Hong A/S

Ostergade 3
4270 Hong
Denmark

with production facility(ies)
42940

is authorized to label the following products with the
TÜV Mark P.

Please see notes overleaf.

Product: Pflegebetten
Patient Care Beds

Model: Electrically adjustable care bed
Type: Etude Basic / Etude Classic

Parameters: see enclosure

The compliance of the above-mentioned product with the following essential requirements was tested on a voluntary basis:

EN 1970:2000
Only the relevant parts
concerning the home-care
were quoted from the
EN 60601-2-38+A1:2000.

EN 60601-2-38/A1:2000
(with regard to)
Trendelenburg positioning and
Anti-Trendelenburg positioning
are not adjustable.

Report No: 70030717 , E - (Z1 02 11 42940 008)

Released with the above certificate number by the
certification body of TÜV PRODUCT SERVICE GMBH.



Department: HANMED5 / ime
Date: 20-Nov-2002

TÜV PRODUCT SERVICE GMBH · Zertifizierstelle · Ridlerstrasse 65 · D-80339 München