



Product Service

Choose certainty.  
Add value.

## Technical Report No. 71349010

Revision: 01

dated 2009-04-20

Client: Invacare EC-Hong A/S  
Ostergade 3  
4270 HONG  
DENMARK

Manufacturing place: Invacare EC-Hong A/S  
Ostergade 3  
4270 HONG  
DENMARK

Test object: Line side rail

Test specifications: FDIS IEC 60601-2-52:2009, section 201.9.8.3.3.102

Purpose of examination: Testing according to the test specifications.

Test result: The test subject was found to be in compliance with the test specifications.

This technical report may only be quoted in full. Any use for advertising purposes must be granted in writing. This report is the result of a single examination of the object in question and is not generally applicable evaluation of the quality of other products in regular production.