



Product Service

## Technical Report No. 713001418

Revision: 1

dated 2012-02-20

Choose certainty.  
Add value.

Client: Invacare France Operations SAS  
Route de St. Roch  
37230 Fondettes

FRANCE

Manufacturing place: Invacare France Operations SAS  
Route de St. Roch  
37230 Fondettes

FRANCE

Test object: Collapsible side rail for medical beds, type Verso II  
Article number:  
- 1554348 (Verso II 30x40)  
- 1554518 (Verso II 20x50)

Test specifications: EN 60601-2-52:2010, clause 201.9.8.3.3.3

Purpose of examination: Testing according to the test specification.

Test result: The test results show that the presented product is in compliance with the specified requirements.

This technical report may only be quoted in full. Any use for advertising purposes must be granted in writing. This report is the result of a single examination of the object in question and is not generally applicable evaluation of the quality of other products in regular production.

File: TR 713001418 Rev 1.docx  
Rep.-No: 713001418  
Revision: 1  
Page 1 of 3

Project Manager:  
Dirk Wexlberger  
Date: 2012-02-20

Phone: + 49 511 9663-863  
Fax: + 49 511 9663-839

E-Mail: [dirk.wexlberger@tuev-sued.de](mailto:dirk.wexlberger@tuev-sued.de)

**TUV**<sup>®</sup>

TÜV SÜD Product Service GmbH

Hanover Branch  
Masurenweg 1-3  
30163 Hanover  
Germany